

Calibration certificate

No. 2026-114

This certificate records the annual calibration of the thickness gauge below. It is written as PDF/A-3U so that it can be read, read the same way, and searched for as long as it has to be kept - which for a calibration record is the lifetime of the instrument plus five years.

Property	Value
Instrument	Eddy current gauge, type N
Serial number	N-4471
Calibrated on	30 July 2026
Reference standard	Foil set, ISO 2360, traceable
Deviation found	0.4 um
Next calibration due	30 July 2027

Signed

Calibration laboratory, Bochum

